

**CERTIFICATE OF CONFORMANCE
EUROPEAN COMMUNITY
COUNCIL DIRECTIVE 2004/108/EC**

Date of Issue: November 2015

Issued By: Retlif Testing Laboratories
795 Marconi Avenue
Ronkonkoma, NY 11779

Issued To: Zengar Institute Inc.
520 Island Road
Victoria, BC Canada V8S-2T6

Reference: Retlif Report Number R-15732-1

Retlif Testing Laboratories hereby acknowledges that compliance testing in accordance with the below listed standards was performed on a representative sample of the equipment listed below. Retlif Testing Laboratories further acknowledges that the test sample listed below was found to be in compliance with these standards.

This certificate is hereby issued to the above named grantee and is valid only for the equipment identified below.

Manufacturer: Zengar Institute Inc.
520 Island Road
Victoria, BC Canada V8S-2T6

Equipment Tested: zAmp

Serial Number: 2985

Brand Name: NeurOptimal

Product Type: Information Technology Equipment (ITE)

- Notes:
- 1) See attached report R-15732-1 for details and/or conditions pertaining to this certificate.
 - 2) Conforms to the emissions requirements of:
 - EN 55022:2010 Conducted Emissions, Mains Ports, Class B
 - EN 55022:2010 Radiated Emissions, Class B
 - EN 61000-3-2:2006/A1:2009/A2:2009 Power Frequency Harmonics
 - EN 61000-3-3:2008 Voltage Fluctuation and Flicker
 - 3) Conforms to the immunity requirements of EN 55024:2010:
 - IEC 61000-4-2:2008 Electrostatic Discharge
 - IEC 61000-4-3:2006/A1:2007/A2:2010 Radiated Immunity
 - IEC 61000-4-4:2004 EFT/Burst, Power Ports and I/O Ports
 - IEC 61000-4-5:2005 Surge Immunity, Power Ports
 - IEC 61000-4-6:2008 Conducted Immunity, Power Ports and I/O Ports
 - IEC 61000-4-11:2004 Voltage Dips and Interrupts